

PCTWORLD INTELLECTUAL PROPERTY
ORGANIZATION

INTERNATIONAL APPLICATION PUBLISHED UNDER

(51) International Patent Classification⁶ :**A61B****A2**

(11)

WO 9608992A2

(43) International Publication Date:

28 March 1996 (28.03.96)(21) International Application Number: **PCT/US95/11686**(22) International Filing Date: **13 September 1995 (13.09.95)**

(30) Priority Data:

110973**14 September 1994 (14.09.94) IL**

(71) Applicants (for all designated States except US): **RAMOT UNIVERSITY AUTHORITY FOR APPLIED RESEARCH & INDUSTRIAL DEVELOPMENT LTD. [IL/IL]; 32 Haim Levanon Street, Ramat Aviv, 69975 Tel Aviv (IL). AK-SELROD, Solange [IL/IL]; Rahvat Ilan 18, 51905 Ramat Ilan (IL). SHOSHAN, Herbert, Z. [US/IL]; Agur 6, 71908 Makabim (IL).**

(72) Inventor; and

(75) Inventor/Applicant (for US only): **KESELBRENER, Laurence [IL/IL]; Beit Tzuri 28, 69122 Tel Aviv (IL).**

(74) Agent: **FRIEDMAN, Mark, M.; c/o Sheinbein, Robert, 2940 Birchtree Lane, Silver Spring, MD 20906 (US).**

(81) Designated States: **AM, AT, AU, BB, BG, BR, BY, CA, CH, CN, CZ, DE, DK, EE, ES, FI, GB, GE, HU, JP, KE, KG, KP, KR, KZ, LK, LR, LT, LU, LV, MD, MG, MN, MW, MX, NO, NZ, PL, PT, RO, RU, SD, SE, SG, SI, SK, TJ, TT, UA, UG, US, UZ, VN, European patent (AT, BE, CH, DE, DK, ES, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE), OAPI patent (BF, BJ, CF, CG, CI, CM, GA, GN, ML, MR, NE, SN, TD, TG), ARIPO patent (KE, MW, SD, SZ, UG).**

Published

Without international search report and to be republished upon receipt of that report.

(54) Title: **APPARATUS AND METHOD FOR TIME DEPENDENT POWER SPECTRUM ANALYSIS OF PHYSIOLOGICAL SIGNALS**

(57) Abstract

An apparatus for time dependent power spectrum analysis of a physiological signal modulated by the autonomic nervous system. The apparatus includes a sensor for picking up a physiological signal modulated by the autonomic nervous system and a frequency selection apparatus for selecting at least one frequency inherent to the signal. The apparatus further includes a selective windowed time-frequency analysis processor for determining the power spectrum of the physiological signal within a window along the signal for the at least one frequency and an output apparatus for providing information associated with the functioning of the autonomic nervous system as provided by the power spectrum of the physiological signal.

FOR THE PURPOSES OF INFORMATION ONLY

Codes used to identify States party to the PCT on the front pages of pamphlets publishing international applications under the PCT.

AT	Austria	GB	United Kingdom	MR	Mauritania
AU	Australia	GE	Georgia	MW	Malawi
BB	Barbados	GN	Guinea	NE	Niger
BE	Belgium	GR	Greece	NL	Netherlands
BF	Burkina Faso	HU	Hungary	NO	Norway
BG	Bulgaria	IE	Ireland	NZ	New Zealand
BJ	Benin	IT	Italy	PL	Poland
BR	Brazil	JP	Japan	PT	Portugal
BY	Belarus	KE	Kenya	RO	Romania
CA	Canada	KG	Kyrgyzstan	RU	Russian Federation
CF	Central African Republic	KP	Democratic People's Republic of Korea	SD	Sudan
CG	Congo	KR	Republic of Korea	SE	Sweden
CH	Switzerland	KZ	Kazakhstan	SI	Slovenia
CI	Côte d'Ivoire	LI	Liechtenstein	SK	Slovakia
CM	Cameroon	LK	Sri Lanka	SN	Senegal
CN	China	LU	Luxembourg	TD	Chad
CS	Czechoslovakia	LV	Latvia	TG	Togo
CZ	Czech Republic	MC	Monaco	TJ	Tajikistan
DE	Germany	MD	Republic of Moldova	TT	Trinidad and Tobago
DK	Denmark	MG	Madagascar	UA	Ukraine
ES	Spain	ML	Mali	US	United States of America
FI	Finland	MN	Mongolia	UZ	Uzbekistan
FR	France			VN	Viet Nam
GA	Gabon				

APPARATUS AND METHOD FOR TIME DEPENDENT POWER SPECTRUM ANALYSIS OF PHYSIOLOGICAL SIGNALS

FIELD AND BACKGROUND OF THE INVENTION

The present invention relates to apparatus and method for time
5 dependent power spectrum analysis of physiological signals in general and
in particular to time dependent power spectrum analysis of cardio-
respiratory physiological signals modulated by the Autonomic Nervous
System (ANS).

It is well known that many physiological signals are modulated by
10 the Autonomic Nervous System (ANS). Such physiological signals include
cardio-respiratory signals including, respiration, ECG, heart rate (HR),
blood pressure (BP), blood flow, vascular resistance, cardiac volume,
cardiac cross section, cardiac contractility, peripheral resistance, and the
like. Other physiological signals which are not modulated by the ANS
15 include EEG signals, EMG signal, ECoG signals, and the like.

It is also well known that perturbations and/or transient changes
which affect the functioning of the ANS affect the physiological signals
and vice versa. Common perturbations used in the analysis of autonomic
control include changing of posture, tilt, pharmaceutical interventions, deep
20 breaths, vagal maneuvers, hand grip, and others.

Generally speaking, the power spectrum of physiological signals in
humans modulated by the ANS can be divided into two frequency ranges:
the Low Frequency (LF) range below 0.15 Hz and the High Frequency
(HF) range above 0.15 Hz displaying a peak at about 0.2 Hz for adults and
25 a peak at about 0.4 Hz for children. The HF range is mediated by the fast
reacting parasympathetic nervous system while the LF range is mediated
by both the parasympathetic nervous system and the slower reacting
sympathetic nervous system.

Standard spectral analysis by Fourier transform or variations of
30 Auto-Regressive Models have been extensively applied in an attempt to

evaluate physiological signals quantitatively under steady state conditions. However, these approaches limit the analysis to rest or restabilization conditions and are not suitable for fast or transient responses which negate the assumption of stationarity.

5 More recently, approaches have been developed which overcome the stationarity limitations of standard spectral analysis. These approaches can be classified into two main groups: time frequency distributions and time dependent models. Time frequency distributions include the Short Time Fourier Transform (STFT), distributions belonging to the Cohen's class
10 such as the Wigner-Ville Distribution (WVD), Exponential Distribution (ED), and the like. Time dependent models are based on Auto Regressive (AR) or Auto Regressive Moving Average (ARMA) modeling. The disadvantages of the above described approaches include the compromise between frequency resolution and quasi-stationarity for the STFT, the
15 smoothing required by WVD to remove interference terms, the importance of the empirically chosen forgetting factor and the model order of the time dependent AR and ARMA models.

A still more recent advance in the analysis of time dependent signals is described in a paper entitled "The Wavelet Transform, Time Frequency
20 Localization and Signal Analysis" by I. Daubechies, IEEE Transactions on Information Theory, Vol. 36. No. 5, September 1990 which is incorporated herein be reference as if set forth fully herein. However, up to the present time, this approach has been limited to EEG signals as described in an article entitled "EEG Paroxysmic Activity Detected by
25 Neural Networks after Wavelet Transform Analysis" by Clochon et al. in the European Symposium on Artificial Neural Networks (ESANN) 1993 Proceedings, pg 145-150.

There is therefore a need for apparatus and method for time dependent power spectrum analysis of physiological signals in general and
30 of cardio-respiratory physiological signals modulated by the Autonomic

Nervous System (ANS) in particular which overcome the above-mentioned deficiencies.

SUMMARY OF THE INVENTION

The present invention is for an apparatus and method for time
5 dependent power spectrum analysis of physiological signals in general and
of cardio-respiratory physiological signals modulated by the Autonomic
Nervous System (ANS) in particular.

Hence, there is provided according to a first aspect of the present
invention, an apparatus for time dependent power spectrum analysis of a
10 physiological signal modulated by the autonomic nervous system, the
apparatus comprising: (a) a sensor for picking up a physiological signal
modulated by the autonomic nervous system; (b) frequency selection
apparatus for selecting at least one frequency inherent to the signal; (c) a
selective windowed time-frequency analysis processor for determining the
15 power spectrum of the physiological signal within a window along the
signal for the at least one frequency; and (d) output apparatus for providing
information associated with the functioning of the autonomic nervous
system as provided by the power spectrum of the physiological signal.

According to a further feature of the present invention, the sensor
20 is selected from one of the following: an ECG recorder, a respiratory
monitor, a blood pressure transducer, a Doppler flow meter, a tachometer,
a finger plethysmograph, a skin resistance galvanometer or any other
cardiovascular monitoring equipment.

According to a still further feature of the present invention, the
25 window has an aperture selected from one of the following: a rectangular
aperture, a Hamming aperture, a Hanning aperture, a Blackman aperture,
a Gaussian window, a Lorentzian window, a sinc window, any power of
a sine window and any power of a cosine window, or any derivative of
these windows.

According to a yet still further feature of the present invention, the apparatus further comprising frequency selection apparatus for determining the frequency resolution of the apparatus.

According to a yet still further feature of the present invention, the
5 apparatus further comprising timing selection apparatus for determining the time resolution of the apparatus.

According to a yet still further feature of the present invention, the apparatus further comprising a physiological signal selector for selecting the type of physiological signal picked up the sensor.

10 According to a yet still further feature of the present invention, the apparatus further comprising a type of perturbation selector for selecting the type of perturbation to be applied to a subject under investigation.

According to a yet still further feature of the present invention, the apparatus further comprising frequency selection apparatus for determining
15 the range of frequency outputted on the output apparatus.

According to a yet still further feature of the present invention, the apparatus further comprising timing selection apparatus for determining the onset and termination of the time-dependent power spectrum of the physiological signal outputted on the output apparatus.

20 According to a yet still further feature of the present invention, the apparatus further comprising a detrending filter deployed between the sensor and the selective windowed time-frequency analysis processor.

According to a yet still further feature of the present invention, the selective windowed time-frequency analysis processor is a Wavelet
25 processor.

According to a yet still further feature of the present invention, the apparatus further comprising a prototype function selector for selecting a prototype function applied by the Wavelet processor.

According to a yet still further feature of the present invention, the

apparatus further comprising a scale parameters selector for selecting scale parameters applied by the Wavelet processor.

According to a yet still further feature of the present invention, the apparatus further comprising a shift parameters selector for selecting shift
5 parameters applied by the Wavelet processor.

According to a yet still further feature of the present invention, the selective windowed time-frequency analysis processor is a selective discrete spectral transform algorithm processor.

According to a yet still further feature of the present invention, the
10 apparatus further comprising a spectral transform selector for selecting a transform selected from one of the following: the Fourier transform, the Haar transform, the Hartley transform, the sine transform, the cosine transform, and the Hadamard transform.

According to a yet still further feature of the present invention, the
15 apparatus further comprising a window duration selector for selecting the duration of the window.

According to a yet still further feature of the present invention, the apparatus further comprising a window aperture selector for selecting the aperture of the window.

20 According to a yet still further feature of the present invention, the apparatus further comprising a number of points selector for selecting the number of data points within the window.

According to a yet still further feature of the present invention, the apparatus further comprising a decimation technique selector for selecting
25 the decimation technique applied by the selective discrete spectral transform analysis processor.

There is also provided according to a second aspect of the present invention, a method of time dependent power spectrum analysis of a physiological signal modulated by the autonomic nervous system, the
30 method comprising the steps of: (a) picking up a physiological signal

modulated by the autonomic nervous system; (b) selecting at least one frequency inherent to the signal; (c) determining the power spectrum of the physiological signal within a window along the signal for the at least one frequency; and (d) providing information associated with the functioning
5 of the autonomic nervous system as provided by the power spectrum of the physiological signal.

There is further provided according to a third aspect of the present invention, an apparatus for time dependent power spectrum analysis of a physiological signal, the apparatus comprising: (a) a sensor for picking up
10 a physiological signal; (b) a frequency selector for selecting at least one frequency inherent to the signal; (c) window generating apparatus for generating a window along the signal, the duration of the window being substantially inversely related to the at least one frequency; (d) decimation apparatus for providing data points within the window; (e) spectral
15 transform apparatus for determining the power spectrum of the signal within the window for the at least one frequency; and (f) output apparatus for providing information associated with time dependent power spectrum of the signal.

According to a further feature of the present invention, the sensor
20 is selected from one of the following: an ECG recorder, a respiratory monitor, a finger plethysmograph, a Doppler flow meter, a tachometer, a blood pressure transducer, a skin resistance galvanometer, an EEG monitor and an EMG monitor.

According to a still further feature of the present invention, the
25 window having an aperture selected from one of the following: a rectangular aperture, a Hamming aperture, a Hanning aperture, a Blackman aperture, a Gaussian window, a Lorentzian window, a sinc window, any power of a sine window and any power of a cosine window or any derivative of these windows.

30 According to a yet still further feature of the present invention, the

apparatus further comprising frequency selection apparatus for determining the frequency resolution of the apparatus.

According to a yet still further feature of the present invention, the apparatus further comprising timing selection apparatus for determining the
5 time resolution of the apparatus.

According to a yet still further feature of the present invention, the apparatus further comprising a physiological signal selector for selecting the type of physiological signal picked up the signal.

According to a yet still further feature of the present invention, the
10 apparatus further comprising a type of perturbation selector for selecting the type of perturbation to be applied to a subject under investigation.

According to a yet still further feature of the present invention, the apparatus further comprising frequency selection apparatus for determining the range of frequency outputted on the output apparatus.

15 According to a yet still further feature of the present invention, the apparatus further comprising timing selection apparatus for determining the onset and termination of the time dependent power spectrum of the physiological signal outputted on the output apparatus.

According to a yet still further feature of the present invention, the
20 apparatus further comprising a detrending filter deployed between the sensor and the selective windowed time-frequency analysis processor.

According to a yet still further feature of the present invention, the apparatus further comprising a spectral transform selector for selecting a transform selected from one of the following: the Fourier transform, the
25 Haar transform, the Hartley transform, the sine transform, the cosine transform, and the Hadamard transform.

According to a yet still further feature of the present invention, the apparatus further comprising a window duration selector for selecting the duration of the window.

30 According to a yet still further feature of the present invention, the

apparatus further comprising a window aperture selector for selecting the aperture of the window.

According to a yet still further feature of the present invention, the apparatus further comprising a number of points selector for selecting the
5 number of data points within the window.

According to a yet still further feature of the present invention, the apparatus further comprising a decimation technique selector for selecting the decimation technique applied by the selective discrete spectral transform analysis processor.

10 There is still further provided according to a fourth aspect of the present invention, a method of time dependent power spectrum analysis of a physiological signal, the method comprising the steps of: (a) picking up a physiological signal; (b) selecting at least one frequency inherent to the signal; (c) generating a window along the signal, the duration of the
15 window being substantially inversely related to the at least one frequency; (d) providing data points within the window; (e) determining the power spectrum of the signal within the window for the at least one frequency; and (f) providing information associated with time dependent power spectrum of the signal.

20 BRIEF DESCRIPTION OF THE DRAWINGS

The invention is herein described, by way of example only, with reference to the accompanying drawings, wherein:

FIG. 1 is a block diagram of a preferred embodiment of an apparatus, constructed and operative according to the teachings of the
25 present invention, employing a selective windowed time-frequency analysis processor for time dependent power spectrum analysis of physiological signals;

FIG. 2 is a representative trace of the heart rate detected from an ECG signal picked up by an ECG recorder of a subject who undergoes a

perturbation in the form of a change in posture from a supine position to a standing position;

FIG. 3 is a block diagram of the selective windowed time-frequency analysis processor of Figure 1 realized as a wavelet processor;

5 FIG. 4 is a block diagram of the selective windowed time-frequency analysis processor of Figure 1 realized as a selective discrete spectral transform processor;

FIGS. 5a-5c are oscilloscope type time dependent graphs of the LF power, the HF power and the LF/HF ratio of the heart rate fluctuations of
10 Figure 2, respectively; and

FIG. 6 is a 2D time dependent power spectrum contour map of the 3D time dependent power spectrum graph shown in Figure 1.

DESCRIPTION OF THE PREFERRED EMBODIMENTS

The present invention is of an apparatus and method for time
15 dependent power spectrum analysis of physiological signals in general and of cardio-respiratory physiological signals modulated by the Autonomic Nervous System (ANS) in particular.

The principles and operation of the apparatus and method according to the present invention may be better understood with reference to the
20 drawings and the accompanying description.

Referring now to the drawings, Figure 1 shows an apparatus, generally designated 100, constructed and operative according to the teachings of the present invention, for analyzing the time dependent power spectrum of physiological signals in general and cardio-respiratory
25 physiological signals modulated by the autonomic nervous system (ANS) in particular.

Generally speaking, apparatus 100 performs time dependent power spectrum analysis of a stationary or a non-stationary, mono-component or multi-component physiological signal picked-up by a sensor 102 adapted

for detecting the physiological signal. The physiological signals include cardio-respiratory signals including, respiration, ECG, heart rate (HR), blood pressure (BP), blood flow, vascular resistance, cardiac volume, cardiac cross section, cardiac contractility, peripheral resistance, and the
5 like. Other physiological signals which are not modulated by the ANS include EEG signals, EMG signal, ECoG signals, and the like. Hence, sensor 102 can be an ECG recorder, a respiratory monitor, a blood pressure transducer, a Doppler flow meter, a tachometer, a finger plethysmograph, a skin resistance galvanometer, a skin resistance galvanometer,
10 cardiovascular monitoring equipment, and the like.

For the sake of exposition, Figure 2 depicts a representative trace of the heart rate as detected from an ECG signal as picked up by an ECG recorder of a human subject undergoing a perturbation in the form of a change in posture (CP) at about time $t = 1225$ seconds to be analyzed by
15 apparatus 100. As is well known, a change in posture (CP) induces an increase in the heart rate of the subject. The increase in heart rate is caused by changes in the modulation of the heart rate by both branches of the autonomic nervous system. First, by an attenuation in the modulation exerted by the fast reacting parasympathetic nervous system. And second,
20 by the strengthening in the modulation of the heart rate exerted by the slower reacting sympathetic nervous system.

The attenuation in the modulation of the heart rate exerted by the parasympathetic nervous system is typically manifested by a sudden overshoot in the heart rate just after the perturbation and then a gradual
25 increase in heart rate such that the average heart rate is greater after the perturbation than before the perturbation. The strengthening in the modulation of the heart rate exerted by the slower reacting sympathetic nervous system is manifested by the gradual increase in the average heart rate from about 1300 seconds to about 1500 seconds.

It is a particular feature of the present invention that apparatus 100 includes a selective windowed time-frequency analysis (SWTFA) processor, generally designated 104, for providing the time-dependent power spectrum analysis of the physiological signal picked up by sensor 102. SWTFA
5 processor 104 can be realized by several implementations including, but not limited to, a Wavelet processor 106 as described in greater detail hereinbelow with reference to Figure 3, a Selective Discrete Spectral Transform Analysis processor 108 as described in greater detail hereinbelow with reference to Figure 4, and the like.

10 In both cases, SWTFA processor 104 includes a window generator 110 for generating a series of windows along the signal within which the power spectrum of the frequencies under investigation is to be analyzed. Furthermore, SWTFA processor 104 includes a power spectrum determination apparatus 112 for determining the power spectrum for a
15 particular frequency within each window provided by window generator 110. Depending on the particular implementation of SWTFA processor 104, the aperture of the windows can be, but are not limited to, a rectangular aperture, a Hamming aperture, a Blackman aperture, a Gaussian window, a Lorentzian window, a sinc window, any power of a sine
20 window, any power of a cosine window, any derivative of these windows, and the like.

It is a particular feature of SWTFA processor 104 that the duration of the windows is generally inversely proportional to the frequency under investigation. Hence, low frequencies are investigated using long time
25 windows while high frequencies are investigated using short time windows. It is a further feature of SWTFA processor 104 that a physiological signal can be analyzed at a wide range of both frequency resolutions and time resolutions. Typically, SWTFA processor 104 is capable of a frequency resolution in the order of 0.001 Hz at the low frequency end of the
30 spectrum for, say, a duration of 30 minutes or even lower for a longer time

duration. Furthermore, SWTFA processor 104 is capable of a time resolution which can reach 1 second for frequencies of around 2 Hz. The time and frequency resolutions reach intermediate values around the center of the time-frequency plane.

5 To regulate the function of SWTFA processor 104, apparatus 100 includes an operator interface 116 enabling an operator to select the type of physiological signal to be analyzed, the type of perturbation, and other operating parameters as will be described hereinbelow. Hence, operator interface 116 includes a physiological signal selector 118 for selecting the
10 type of physiological signal to be analyzed and a type of perturbation selector 120 for selecting the type of perturbation to be applied to a subject under investigation.

The frequencies or ranges of frequencies under investigation are preferably selected by a user using frequency selection apparatus 122.
15 Typically, one or more frequencies can be selected for investigation from a range from about 0 Hz and about 20 Hz for the analysis of the autonomic nervous systems of a wide range of mammals. The frequencies under investigation are determined by assigning values to a first frequency register 124, a last frequency register 126 and a frequency interval register
20 128 such that SWTFA processor 104 processes a sequence of frequencies starting at the first frequency stored in first frequency register 124 and increasing in steps determined by frequency interval register 128 up to the last frequency stored in last frequency register 126. Furthermore, frequency selection apparatus 122 preferably includes one or more
25 frequency range selectors 130 and 132 for determining the display of their corresponding power spectrum. Typically, frequency range selectors 130 and 132 are used for selecting displays of LF power and HF power.

Operator interface 116 also includes timing selection apparatus 134 which includes a timer 136 for determining the time resolution of the
30 analysis of the physiological signal and several pairs of START and STOP

timers as described hereinbelow. It should be noted that timer 136 determines the degree of overlapping between consecutive windows. A first pair of timers 138 is preferably dedicated to controlling the onset and termination of the display of the LF power spectrum of the autonomic control modulating the physiological signal. A second pair of timers 140 is preferably dedicated to controlling the onset and termination of the display of the HF power spectrum of the autonomic control moderating the physiological signal. Other timers 142 can be assigned by the operator to determine time intervals during which other parameters can be displayed or calculated.

SWTFA processor 104 is typically connected to sensor 102 via an amplifier 142, an anti aliasing filter 144, and an A/D converter 148. Other devices may be required to provide suitable data to SWTFA processor 104 depending on the type of signal being picked up by sensor 102. For example, in the case of sensor 102 picking up heartbeats, then an R wave detector is required to provide suitable data to SWTFA processor 104. The additional devices required to be provide suitable data to SWTFA processor 104 can be connected through selection of the type of physiological sensor to be picked up by sensor 102 using physiological signal selector 118.

It is a further feature of the present invention that apparatus 100 preferably includes a detrending filter 150 deployed between A/D converter 146 and SWTFA processor 104 such that sudden changes of amplitude in the physiological signal after the perturbation do not affect the natural high frequency components of the physiological signal. The parameters of detrending filter 150 can be determined by the settings of physiological signal selector 118 and type of perturbation selector 120. Alternatively, the parameters of detrending filter 150 can be set by the operator using detrending filter control 152.

Apparatus 100 further includes a display driver 154 for providing output on an output apparatus realized as a display monitor 156 and a

normalization apparatus 158 for normalizing the output on display monitor 156. Output apparatus can also include a printer. The displays can include 3D graphic displays, 2D contour map displays, time dependent oscilloscope type displays, and the like depending on the type of information required
5 by an operator. In particular, display driver 154 can display a 3D time dependent power spectrum graph 160 where the x-axis 162 of graph 160 depicts frequency in Hz, the y-axis 164 of graph 160 depicts time in seconds and the z-axis 166 of graph 160 depicts the power spectrum in units associated with the type of physiological signal being picked up by
10 sensor 102. For example, z-axis 166 can have units of beats per minute² Hz⁻¹ (BPM² Hz⁻¹) when the physiological signal measured is heart beat, mmHg² Hz⁻¹ when the physiological signal measured is blood pressure, (mlsec⁻¹)² Hz⁻¹ when the physiological signal measured is blood flow, and the like.

15 Furthermore, apparatus 100 includes a metrics processor 168 for providing one or more metrics depending on the type of information required by an operator. Metrics processor 168 typically receives input from a number of sources including, but not limited to, physiological signal selector 118, type of perturbation selector 120, frequency selection
20 apparatus 122, timing selection apparatus 134, and the like. Hence, metrics processor 168 can provide the baseline LF power before perturbation, the LF power during or after perturbation, maximal and/or minimal activity during perturbation and the ratio therebetween. In a similar fashion, metrics processor 168 can provide the baseline HF power before
25 perturbation, the HF power during or after perturbation, maximal and/or minimal activity during perturbation and the ratio therebetween. Still further, metrics processor 168 can provide the LF/HF ratio before perturbation to the LF/HF ratio after perturbation.

With reference now to Figure 3, the block diagram depicts the
30 preferred embodiment of SWTFA processor 104 realized as Wavelet

processor 106 including window generator 106 additionally designated 170 and power spectrum determination apparatus 112 additionally designated 172. For further information on wavelet processing, reference is made to the aforementioned paper entitled "The Wavelet Transform, Time
5 Frequency Localization and Signal Analysis" by I. Daubechies, IEEE Transactions on Information Theory, Vol. 36. No. 5, September 1990 which is incorporated herein by reference as if set forth fully herein.

Broadly speaking, the aperture, duration and the time resolution between consecutive windows provided by window generator 170 are
10 defined by three parameters: a prototype function $h(t)$, a scale parameter "a" and a shift parameter "b" according to the Wavelet transform:

$$\int h_{ab}(t) f(t) \cdot dt \text{ where } a \in \mathbb{R}^+, b \in \mathbb{R}$$

As well known in the art, for a large scale parameter value, the prototype function is stretched such that the prototype Wavelet acts as a low
15 frequency function while, for a small scale parameter value, the prototype function is contracted such that the Wavelet function acts a high frequency function. Hence, depending on the value assigned to scaling parameter "a", the Wavelet function dilates or contracts in time, causing the corresponding contraction or dilation in the frequency domain. Thus, the Wavelet
20 transform provides a flexible time-frequency resolution and analyzes higher frequencies with better time resolution but poorer frequency resolution than lower frequencies.

Preferably, apparatus 100 is configured such that the settings of physiological signal selector 118 and type of perturbation selector 120
25 determine the selections of prototype function $h(t)$, effective scale parameters "a" and effective shift parameters "b", thereby obviating the need for direct operator intervention. Alternatively, apparatus 100 can include a prototype function selector 174 for selection of the prototype function $h(t)$, a scale parameters selector 176 for selection of the scale
30 parameter "a" and a shift parameters selector 178 for selection of shift

parameter "b", thereby enabling operator intervention in the regulation of Wavelet processor 106 according to the type of physiological signal to be analyzed, the type of perturbation, and the like. Depending on the prototype function selected, the aperture of the windows can be, but are not
5 limited to, a rectangular aperture, a Hamming aperture, a Blackman aperture, a Gaussian window, a Lorentzian window, a sinc window, any power of a sine window, any power of a cosine window, any derivative of these windows, and the like.

With reference now to Figure 4, the block diagram depicts the
10 preferred embodiment of SWTFA processor 104 realized as Selective Discrete Spectral Transform Algorithm processor (SDA) 108 including window generator 106 additionally designated 180 and power spectrum determination apparatus 112. In this case, power spectrum determination apparatus 112 includes decimation apparatus 182 for providing a constant
15 number of data points from the windows, a spectral transform apparatus 184 for providing the power spectrum of the frequency within the windows based on the data points and a window correction apparatus 186 for correcting the power spectrum due to distortions rendered by window generator 180.

20 SDA apparatus 108 can employ any one of a number of known transforms for determining the power spectrum of a time dependent physiological signal. Such transforms include the Fourier transform, the Haar transform, the Hartley transform, the sine transform, the cosine transform, the Hadamard transform, and the like. The selection of the
25 spectral transform realized by SDA apparatus 180 is preferably operator controlled by means of a spectral transform selector 188.

As mentioned hereinabove, the duration of windows is preferably inversely related to the frequency under investigation. Depending on the type of signal, the duration of windows typically lies from about 2 periods
30 and about 10 periods of the frequency under investigation. The duration

is preferably user selected by means of a window duration selector 190. Alternatively, the duration of the window can be automatically set according to the SNR or similar metric of the physiological signal.

The windows can have different apertures including, but not limited
5 to, a rectangular aperture, a Hamming aperture, a Hanning aperture, a Blackman aperture, a Gaussian window, a Lorentzian window, a sinc window, any power of a sine window, any power of a cosine window, any derivative of these windows, and the like. The aperture of the window is preferably user selected by means of a window aperture selector 192. In
10 the case that spectral transform apparatus 184 employs the Fourier transform and a rectangular window is employed to ensure the highest possible frequency resolution by minimizing side lobes, then window correction apparatus 186 is required to correct the obtained power spectra by dividing by the corresponding sinc function. For other windows and
15 transforms, the correction required depends on the type of window and the mathematical rules of the transform.

Decimation apparatus 182 preferably provides data points by employing a low pass filter and undersampling technique including, but not limited to, moving average. Typically, decimation apparatus 182 provides
20 the same number of data points irrespective of the duration of the windows so as not to generate artifacts or normalization problems. The number of data points determined in a window is preferably operator selected by means of a number of data points selector 194. Furthermore, SDA apparatus 108 preferably includes a decimation technique selector 196 for
25 selecting a technique for decimation of a window.

In a similar manner to Wavelet processor 106, SDA apparatus 108 can be configured such that settings of physiological signal selector 118 and type of perturbation selector 118 determines: the settings of spectral transform selector 188, window duration selector 190, window aperture

selector 192, number of data points selector 194 and decimation technique selector 196.

The operation of apparatus 100 is now described with reference to the time dependent power spectrum analysis of the physiological signal 5 depicted in Figure 2 of the heart rate of a human subject who has undergone a perturbation in the form of a change of posture at time $t=1225$ seconds. Hence, an operator selects an appropriate sensor 102, in this case an ECG recorder, for detecting the heart rate of a subject.

Then the operator employs the above-mentioned selectors in the 10 following fashion. First, the operator employs physiological signal selector 118 for selecting the type of physiological signal to be analyzed and perturbation selector 120 for selecting the type of perturbation to be applied to a subject under investigation. These selections can cause connection of an R wave detector if required and the setting of SWTFA processor 104 15 as described hereinabove. And second, the operator employs frequency selection apparatus 122 to determine the frequencies under investigation by assigning values to registers 124, 126 and 128.

In the case that SWTFA processor 104 is realized as Wavelet processor 106, the operator can employ prototype function selector 174, 20 scale parameters selector 176 and shift parameters selector 178 for setting the prototype function, the scale parameters and the shift parameters of Wavelet processor 106, respectively. While in the case that SWTFA processor 104 is realized as SDA processor 108, the operator employs spectral transform selector 188, window duration selector 190, window 25 aperture selector 192, number of data points selector 194 and decimation technique selector 196 to determine the settings of SDA apparatus 108.

Apparatus 100, and in particular SWTFA processor 104, analyzes the physiological signal in the following manner. Window generator 110 generates a window toward the onset of the pick-up of the physiological 30 signal for the first frequency under investigation according to the selection

of the operator. Power spectrum determination apparatus **112** determines the power spectrum of the frequency within the window and provides the output to both display driver **154** and metrics processor **168**. In the same manner, SWTFA processor **104** analyzes the window for other frequencies under investigation according to the selection of the operator such that apparatus **100** provides a real time, time dependent, power spectrum analysis of the physiological signal. It should be noted as described hereinabove that the size of the windows varies as a function of the frequency under investigation. After determining the power spectrum within the window for each frequency under investigation, SWTFA processor **104** repeats the method for the second window deployed forward along the physiological signal according to the time resolution and frequency resolution set by the operator for the next frequency and so on. Alternatively, SWTFA processor **104** can process all the time signal for each frequency under investigation.

With particular reference to SDA processor **108**, window generator **180** generates a first window starting at the onset of the pick-up of the physiological signal for the first frequency under investigation according to the selection of the operator, decimation apparatus **182** decimates the data points in the window to provide the predetermined number of data points, spectrum transform apparatus **184** determines the power spectrum of the frequency during the first window and provides the output to window correction apparatus **186** for correction of the power spectrum.

With reference now to Figures 5a-5c, apparatus **100** provides the following displays and metrics after the time dependent power spectrum analysis of the physiological signal shown in Figure 2. First, display driver **154** can provide an oscilloscope type time dependent LF power graph **198** according to input received from timing selection apparatus **122** and frequency selection apparatus **134**. In this case, as shown in Figure 5a, time dependent LF power graph **198** is provided for the LF power integral

over the range of from about 0.04 Hz to about 0.12 Hz. Second, display driver 154 can provide an oscilloscope type time dependent HF power graph 200 according to input received from timing selection apparatus 122 and frequency selection apparatus 134. In this case, as shown in Figure 5b, time dependent HF power graph 200 is provided for the HF power spectrum integral over the range of from about 0.24 Hz to about 0.6 Hz. Third, display driver 154 can provide an oscilloscope type graph 202 of the sympatho-vagal balance LF/HF as shown in Figure 5c. Fourth, display driver 154 can display a 2D time dependent power spectrum contour map 204 of time against frequency as shown in Figure 6 in which the contours are representative of the time dependent power spectrum of the fluctuations of the physiological signal by the autonomic nervous system. And fifth, display driver 154 can provide 3D time dependent power spectrum graph 160 as shown in Figure 1. It should be noted that 2D time dependent power spectrum contour map 204 is in fact a top view of 3D time dependent power spectrum graph 160.

Hence, it can be readily appreciated that apparatus 100 provides information regarding the heart rate of the subject as induced by the functioning of the autonomic nervous system due to the change of posture as will now be described in greater detail. First, HF power graph 200 depicts that a sudden attenuation in HF power at the time of perturbation corresponding to the fast response of the parasympathetic nervous system. Second, LF power spectrum graph 198 depicts that there is a strengthening in the LF power after perturbation corresponding to the slower response of sympathetic nervous system. Third, HF power spectrum graph 200, which reflects only the functioning of the parasympathetic nervous system, depicts that HF power is attenuated after the perturbation. And fourth, the LF/HF sympatho-vagal balance ratio graph 202 depicts that the contribution of LF power is greater after the perturbation confirming the known functioning of the ANS during the change of posture.

It can therefore be appreciated that apparatus 100 provides time dependent evaluation of the functioning of the ANS which will be useful in the clinical interpretation of a wide range of clinical conditions including, autonomic changes relative to fainting spells such as vasovagal
5 syncope, tilt test, autonomic failure, autonomic imbalance, diabetic neuropathy, cardiac ischemic changes, effects of drugs interfering with autonomic or ischemic changes, angioplasty effects on autonomic control, early detection of hypertension. Furthermore, apparatus 100 can be used for all kinds of autonomic tests such as: hand grip, cold pressor test, vagal
10 maneuvers such as Oculocardiac Reflex or Diving Response, valsalva maneuver and the like. It can also be appreciated that apparatus 100 can provide time dependent power spectrum analysis of a wide range of other physiological signals in a similar fashion.

While the invention has been described with respect to a limited
15 number of embodiments, it will be appreciated that many variations, modifications and other applications of the invention may be made.

WHAT IS CLAIMED IS:

1. An apparatus for time dependent power spectrum analysis of a physiological signal modulated by the autonomic nervous system, the apparatus comprising:
 - (a) a sensor for picking up a physiological signal modulated by the autonomic nervous system;
 - (b) frequency selection apparatus for selecting at least one frequency inherent to said signal;
 - (c) a selective windowed time-frequency analysis processor for determining the power spectrum of said physiological signal within a window along said signal for said at least one frequency; and
 - (d) output apparatus for providing information associated with the functioning of the autonomic nervous system as provided by the power spectrum of the physiological signal.
2. The apparatus as in claim 1, wherein said sensor is selected from one of the following: an ECG recorder, a respiratory monitor, a blood pressure transducer, a Doppler flow meter, a tachometer, a finger plethysmograph, a skin resistance galvanometer, cardiovascular monitoring equipment.
3. The apparatus as in claim 1, wherein said window has an aperture selected from one of the following: a rectangular aperture, a Hamming aperture, a Hanning aperture, a Blackman aperture, a Gaussian window, a Lorentzian window, a sinc window, any power of a sine window and any power of a cosine window, or any derivative of said windows.

4. The apparatus as in claim 1, further comprising frequency selection apparatus for determining the frequency resolution of the apparatus.
5. The apparatus as in claim 1, further comprising timing selection apparatus for determining the time resolution of the apparatus.
6. The apparatus as in claim 1, further comprising a physiological signal selector for selecting the type of physiological signal picked up said sensor.
7. The apparatus as in claim 1, further comprising a type of perturbation selector for selecting the type of perturbation to be applied to a subject under investigation.
8. The apparatus as in claim 1, further comprising frequency selection apparatus for determining the range of frequency outputted on said output apparatus.
9. The apparatus as in claim 1, further comprising timing selection apparatus for determining the onset and termination of the time dependent power spectrum of the physiological signal outputted on said output apparatus.
10. The apparatus as in claim 1, further comprising a detrending filter deployed between said sensor and said selective windowed time-frequency analysis processor.
11. The apparatus as in claim 1, wherein said selective windowed time-frequency analysis processor is a Wavelet processor.

12. The apparatus as in claim 11, further comprising a prototype function selector for selecting a prototype function applied by said Wavelet processor.
13. The apparatus as in claim 11, further comprising a scale parameters selector for selecting scale parameters applied by said Wavelet processor.
14. The apparatus as in claim 11, further comprising a shift parameters selector for selecting shift parameters applied by said Wavelet processor.
15. The apparatus as in claim 1, wherein said selective windowed time-frequency analysis processor is a selective discrete spectral transform algorithm processor.
16. The apparatus as in claim 15, further comprising a spectral transform selector for selecting a transform selected from one of the following: the Fourier transform, the Haar transform, the Hartley transform, the sine transform, the cosine transform, and the Hadamard transform.
17. The apparatus as in claim 15, further comprising a window duration selector for selecting the duration of said window.
18. The apparatus as in claim 15, further comprising a window aperture selector for selecting the aperture of said window.
19. The apparatus as in claim 15, further comprising a number of points selector for selecting the number of data points within said window.

20. The apparatus as in claim 15, further comprising a decimation technique selector for selecting the decimation technique applied by said selective discrete spectral transform analysis processor.

21. A method of time dependent power spectrum analysis of a physiological signal modulated by the autonomic nervous system, the method comprising the steps of:

- (a) picking up a physiological signal modulated by the autonomic nervous system;
- (b) selecting at least one frequency inherent to the signal;
- (c) determining the power spectrum of the physiological signal within a window along said the for the at least one frequency; and
- (d) providing information associated with the functioning of the autonomic nervous system as provided by the power spectrum of the physiological signal.

22. An apparatus for time dependent power spectrum analysis of a physiological signal, the apparatus comprising:

- (a) a sensor for picking up a physiological signal;
- (b) a frequency selector for selecting at least one frequency inherent to said signal;
- (c) window generating apparatus for generating a window along said signal, the duration of said window being substantially inversely related to said at least one frequency;
- (d) decimation apparatus for providing data points within said window;
- (e) spectral transform apparatus for determining the power spectrum of said signal within said window for said at least one frequency; and

- (f) output apparatus for providing information associated with time dependent power spectrum of said signal.

23. The apparatus as in claim 22, wherein said sensor is selected from one of the following: an ECG recorder, a respiratory monitor, a finger plethysmograph, a Doppler flow meter, a tachometer, a blood pressure transducer, a skin resistance galvanometer, an EEG monitor, an EMG monitor.

24. The apparatus as in claim 22, wherein said window having an aperture selected from one of the following: a rectangular aperture, a Hamming aperture, a Hanning aperture, a Blackman aperture, a Gaussian window, a Lorentzian window, a sinc window, any power of a sine window, and any power of a cosine window, or any derivative of said windows.

25. The apparatus as in claim 22, further comprising frequency selection apparatus for determining the frequency resolution of the apparatus.

26. The apparatus as in claim 22, further comprising timing selection apparatus for determining the time resolution of the apparatus.

27. The apparatus as in claim 22, further comprising a physiological signal selector for selecting the type of physiological signal picked up said signal.

28. The apparatus as in claim 22, further comprising a type of perturbation selector for selecting the type of perturbation to be applied to a subject under investigation.

29. The apparatus as in claim 22, further comprising frequency selection apparatus for determining the range of frequency outputted on said output apparatus.

30. The apparatus as in claim 22, further comprising timing selection apparatus for determining the onset and termination of the time dependent power spectrum of the physiological signal outputted on said output apparatus.

31. The apparatus as in claim 22, further comprising a detrending filter deployed between said sensor and said selective windowed time-frequency analysis processor.

32. The apparatus as in claim 22, further comprising a spectral transform selector for selecting a transform selected from one of the following: the Fourier transform, the Haar transform, the Hartley transform, the sine transform, the cosine transform, and the Hadamard transform.

33. The apparatus as in claim 22, further comprising a window duration selector for selecting the duration of said window.

34. The apparatus as in claim 22, further comprising a window aperture selector for selecting the aperture of said window.

35. The apparatus as in claim 22, further comprising a number of points selector for selecting the number of data points within said window.

36. The apparatus as in claim 22, further comprising a decimation technique selector for selecting the decimation technique applied by said selective discrete spectral transform analysis processor.

37. A method of time dependent power spectrum analysis of a physiological signal, the method comprising the steps of:

- (a) picking up a physiological signal;
- (b) selecting at least one frequency inherent to the signal;
- (c) generating a window along the signal, the duration of the window being substantially inversely related to the at least one frequency;
- (d) providing data points within the window;
- (e) determining the power spectrum of the signal within the window for the at least one frequency; and
- (f) providing information associated with time dependent power spectrum of the signal.

1/6




FIG. 1A

2/6







FIG. 1B

3/6

FIG. 2



4/6



5/6




FIG.5A




FIG.5B





FIG.5C

6/6

FIG. 6





INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(51) International Patent Classification⁶ :

A61B 5/02

A3

(11) International Publication Number:

WO 96/08992

(43) International Publication Date:

28 March 1996 (28.03.96)

(21) International Application Number: PCT/US95/11686

(22) International Filing Date: 13 September 1995 (13.09.95)

(30) Priority Data:

110973

14 September 1994 (14.09.94)

IL

(71) Applicants (for all designated States except US): RAMOT UNIVERSITY AUTHORITY FOR APPLIED RESEARCH & INDUSTRIAL DEVELOPMENT LTD. [IL/IL]; 32 Haim Levanon Street, Ramat Aviv, 69975 Tel Aviv (IL). AK-SELROD, Solange [IL/IL]; Rahvat Ilan 18, 51905 Ramat Ilan (IL). SHOSHAN, Herbert, Z. [US/IL]; Agur 6, 71908 Makabim (IL).

(72) Inventor; and

(75) Inventor/Applicant (for US only): KESELBRENER, Laurence [IL/IL]; Beit Tzuri 28, 69122 Tel Aviv (IL).

(74) Agent: FRIEDMAN, Mark, M.; c/o Sheinbein, Robert, 2940 Birchtree Lane, Silver Spring, MD 20906 (US).

(81) Designated States: AM, AT, AU, BB, BG, BR, BY, CA, CH, CN, CZ, DE, DK, EE, ES, FI, GB, GE, HU, JP, KE, KG, KP, KR, KZ, LK, LR, LT, LU, LV, MD, MG, MN, MW, MX, NO, NZ, PL, PT, RO, RU, SD, SE, SG, SI, SK, TJ, TT, UA, UG, US, UZ, VN, European patent (AT, BE, CH, DE, DK, ES, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE), OAPI patent (BF, BJ, CF, CG, CI, CM, GA, GN, ML, MR, NE, SN, TD, TG), ARIPO patent (KE, MW, SD, SZ, UG).


Published

With international search report.

(88) Date of publication of the international search report:

13 June 1996 (13.06.96)

(54) Title: APPARATUS AND METHOD FOR TIME DEPENDENT POWER SPECTRUM ANALYSIS OF PHYSIOLOGICAL SIGNALS



(57) Abstract

This invention is an apparatus for time dependent power spectrum analysis of a physiological signal modulated by the autonomic nervous system. The apparatus includes a sensor (102) for picking up a physiological signal modulated by the autonomic nervous system and a frequency selection apparatus (122) for selecting at least one frequency inherent to the signal. The apparatus further includes a selective windowed time-frequency analysis processor (104) for determining the power spectrum of the physiological signal within a window along the signal for the at least one frequency, and an output apparatus (154) for providing information associated with the functioning of the autonomic nervous system as provided by the power spectrum of the physiological signal.

FOR THE PURPOSES OF INFORMATION ONLY

Codes used to identify States party to the PCT on the front pages of pamphlets publishing international applications under the PCT.

AT	Austria	GB	United Kingdom	MR	Mauritania
AU	Australia	GE	Georgia	MW	Malawi
BB	Barbados	GN	Guinea	NE	Niger
BE	Belgium	GR	Greece	NL	Netherlands
BF	Burkina Faso	HU	Hungary	NO	Norway
BG	Bulgaria	IE	Ireland	NZ	New Zealand
BJ	Benin	IT	Italy	PL	Poland
BR	Brazil	JP	Japan	PT	Portugal
BY	Belarus	KE	Kenya	RO	Romania
CA	Canada	KG	Kyrgyzstan	RU	Russian Federation
CF	Central African Republic	KP	Democratic People's Republic of Korea	SD	Sudan
CG	Congo	KR	Republic of Korea	SE	Sweden
CH	Switzerland	KZ	Kazakhstan	SI	Slovenia
CI	Côte d'Ivoire	LI	Liechtenstein	SK	Slovakia
CM	Cameroon	LK	Sri Lanka	SN	Senegal
CN	China	LU	Luxembourg	TD	Chad
CS	Czechoslovakia	LV	Latvia	TG	Togo
CZ	Czech Republic	MC	Monaco	TJ	Tajikistan
DE	Germany	MD	Republic of Moldova	TT	Trinidad and Tobago
DK	Denmark	MG	Madagascar	UA	Ukraine
ES	Spain	ML	Mali	US	United States of America
FI	Finland	MN	Mongolia	UZ	Uzbekistan
FR	France			VN	Viet Nam
GA	Gabon				

INTERNATIONAL SEARCH REPORT

International application No.
PCT/US95/11686**A. CLASSIFICATION OF SUBJECT MATTER**

IPC(6) :A61B 5/02

US CL :128/668

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)

U.S. : 128/630, 668, 671, 696, 671, 700, 702; 364/413.02, 413.03, 413.05, 413.06

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

Electronic data base consulted during the international search (name of data base and, where practicable, search terms used)

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X --- Y	US, A, 5,046,504 (ALBERT ET AL.) 10 September 1991, see column 11 lines 21-50.	22-26, 29-37 ----- 1-9, 15-21, 27, 28
Y	US, A, 4,777,960 (BERGER ET AL.) 18 October 1988, see column 4 line 52 to column 8 line 24.	1-9, 15-21, 27, 28
A	US, A, 4,979,110 (ALBRECHT ET AL.) 18 December 1990, see column 5 lines 38-65.	1-37

☐ Further documents are listed in the continuation of Box C.☐ See patent family annex.

* Special categories of cited documents:

- *A* document defining the general state of the art which is not considered to be part of particular relevance
- *E* earlier document published on or after the international filing date
- *L* document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)
- *O* document referring to an oral disclosure, use, exhibition or other means
- *P* document published prior to the international filing date but later than the priority date claimed

T

later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention

X

document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone

Y

document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art

Z

document member of the same patent family

Date of the actual completion of the international search

15 FEBRUARY 1996

Date of mailing of the international search report

04 MAR 1996

Name and mailing address of the ISA/US
Commissioner of Patents and Trademarks
Box PCT
Washington, D.C. 20231

Facsimile No. (703) 305-3230

Authorized officer

ERIC F. WINAKUR

Telephone No. (703) 308-3940

Form PCT/ISA/210 (second sheet)(July 1992)*


This Page Blank (uspto)



INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(51) International Patent Classification ⁶ : A61B 5/02		A3	(11) International Publication Number: WO 96/08992
			(43) International Publication Date: 28 March 1996 (28.03.96)
(21) International Application Number: PCT/US95/11686		(81) Designated States: AM, AT, AU, BB, BG, BR, BY, CA, CH, CN, CZ, DE, DK, EE, ES, FI, GB, GE, HU, JP, KE, KG, KP, KR, KZ, LK, LR, LT, LU, LV, MD, MG, MN, MW, MX, NO, NZ, PL, PT, RO, RU, SD, SE, SG, SI, SK, TJ, TT, UA, UG, US, UZ, VN, European patent (AT, BE, CH, DE, DK, ES, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE), OAPI patent (BF, BJ, CF, CG, CI, CM, GA, GN, ML, MR, NE, SN, TD, TG), ARIPO patent (KE, MW, SD, SZ, UG).	
(22) International Filing Date: 13 September 1995 (13.09.95)			
(30) Priority Data: 110973 14 September 1994 (14.09.94) IL			
(71) Applicants (for all designated States except US): RAMOT UNIVERSITY AUTHORITY FOR APPLIED RESEARCH & INDUSTRIAL DEVELOPMENT LTD. [IL/IL]; 32 Haimi Levanon Street, Ramat Aviv, 69975 Tel Aviv (IL). SHOSHAN, Herbert, Z. [US/IL]; Agur 6, 71908 Makabim (IL).		Published With international search report.	
(72) Inventors; and (75) Inventors/Applicants (for US only): KESELBRENER, Laurence [IL/IL]; Beit Tzuri 28, 69122 Tel Aviv (IL). AKSELROD, Solange [IL/IL]; Rahvat Ilan 18, 51905 Ramat Ilan (IL).		(88) Date of publication of the international search report: 13 June 1996 (13.06.96)	
(74) Agent: FRIEDMAN, Mark, M.; c/o Sheinbein, Robert, 2940 Birchtree Lane, Silver Spring, MD 20906 (US).			

(54) Title: APPARATUS AND METHOD FOR TIME DEPENDENT POWER SPECTRUM ANALYSIS OF PHYSIOLOGICAL SIGNALS



(57) Abstract

This invention is an apparatus for time dependent power spectrum analysis of a physiological signal modulated by the autonomic nervous system. The apparatus includes a sensor (102) for picking up a physiological signal modulated by the autonomic nervous system and a frequency selection apparatus (122) for selecting at least one frequency inherent to the signal. The apparatus further includes a selective windowed time-frequency analysis processor (104) for determining the power spectrum of the physiological signal within a window along the signal for the at least one frequency, and an output apparatus (154) for providing information associated with the functioning of the autonomic nervous system as provided by the power spectrum of the physiological signal.

FOR THE PURPOSES OF INFORMATION ONLY

Codes used to identify States party to the PCT on the front pages of pamphlets publishing international applications under the PCT.

AT	Austria	GB	United Kingdom	MR	Mauritania
AU	Australia	GE	Georgia	MW	Malawi
BB	Barbados	GN	Guinea	NE	Niger
BE	Belgium	GR	Greece	NL	Netherlands
BF	Burkina Faso	HU	Hungary	NO	Norway
BG	Bulgaria	IE	Ireland	NZ	New Zealand
BJ	Benin	IT	Italy	PL	Poland
BR	Brazil	JP	Japan	PT	Portugal
BY	Belarus	KE	Kenya	RO	Romania
CA	Canada	KG	Kyrgystan	RU	Russian Federation
CF	Central African Republic	KP	Democratic People's Republic of Korea	SD	Sudan
CG	Congo	KR	Republic of Korea	SE	Sweden
CH	Switzerland	KZ	Kazakhstan	SI	Slovenia
CI	Côte d'Ivoire	LI	Liechtenstein	SK	Slovakia
CM	Cameroon	LK	Sri Lanka	SN	Senegal
CN	China	LU	Luxembourg	TD	Chad
CS	Czechoslovakia	LV	Latvia	TG	Togo
CZ	Czech Republic	MC	Monaco	TJ	Tajikistan
DE	Germany	MD	Republic of Moldova	TT	Trinidad and Tobago
DK	Denmark	MG	Madagascar	UA	Ukraine
ES	Spain	ML	Mali	US	United States of America
FI	Finland	MN	Mongolia	UZ	Uzbekistan
FR	France			VN	Viet Nam
GA	Gabon				